

***TB 11-6625-2669-35**

DEPARTMENT OF THE ARMY TECHNICAL BULLETIN

CALIBRATION PROCEDURE FOR SIGNAL GENERATOR SG-769/U AND WAVETEK MODEL 111 AND 111B

Headquarters, Department of the Army, Washington, DC
23 July 1976

REPORTING OF ERRORS

You can help improve this publication by calling attention to errors and by recommending improvements and stating your reasons for the recommendations. Your letter or DA Form 2028, Recommended Changes to Publications, should be mailed directly to Commander, U.S. Army Aviation and Missile Command, ATTN: AMSMI-TMDE-EP, Redstone Arsenal, AL 35898-5400. FAX to DSN 788-2313 (commercial 256-842-2313). A reply will be furnished directly to you.

SECTION		Paragraph	Page
	I. IDENTIFICATION AND DESCRIPTION		
	Test instrument identification	1	2
	Calibration data (DA Form 2416).....	2	2
	Calibration description	3	2
	II. EQUIPMENT REQUIREMENTS		
	Equipment required	4	4
	Accessories required	5	4
	III. CALIBRATION PROCESS (SECONDARY OR AN/GSM-256)		
	Preliminary instructions	6	6
	Frequency and stability	7	6
	Voltage-controlled generator linearity	8	9
	Output amplitude and symmetry.....	9	9
	Frequency response.....	10	10
	Alternate frequency response.....	11	11
	Rise time	12	12
	Distortion.....	13	12
	IV. MAINTENANCE CALIBRATION PROCESS		
	Preliminary instructions	14	13

*This bulletin supersedes TB 9-6625-1347-50, 26 September 1969.

TB 11-6625-2669-35

Triangle wave	15	14
Square wave	16	15
Square wave 10 nanoseconds	17	17
Power amplifier	18	17
Sine wave.....	19	18
Frequency dial tracking	20	19
VCG input.....	21	21
Ramp.....	22	23
Range tracking	23	23
Final procedure.....	24	25

**SECTION I
IDENTIFICATION AND DESCRIPTION**

1. Test Instrument Identification. This bulletin provides instructions for the calibration of Signal Generator SG-769M and Wavetek Model 111 and 111B. The manufacturer's instruction manual was used as the prime data source in compiling these instructions. The equipment being calibrated will be referred to as the TI (test instrument) throughout this bulletin.

a. Model Variations. Variations among models are described in text.

b. Time and Technique. The time required for this calibration is approximately 6 hours, using the dc and low frequency technique.

2. Calibration Data (DA Form 2416)

a. Forms, records, and reports required for calibration personnel at all levels are prescribed by TM 38-750. DA Form 2416 must be annotated in accordance with TM 38-750 for each calibration performed. Reportable adjustments are followed by (R) in this procedure. Report only those adjustments made on the DA Form 2416.

3. Calibration Description. TI parameters and performance specifications which pertain to this calibration are listed in table 1.

Table 1. Calibration Description - Test Instrument, Serial Numbers of 017781 or Lower

Test instrument parameters	Performance specifications
Power input requirements*	115 V ac $\pm 10\%$ or 230 V ac $\pm 10\%$ 50 to 400 Hz at approx. 10 w
Frequency	Range: 0.0015 Hz to 1 MHz in 8 ranges Accuracy: $\pm 1\%$ fs from 0.0015 Hz to 100 kHz; $\pm 2\%$ of fs from 100 kHz to 1 MHz
Stability*	Drift less than $\pm 0.05\%$ of setting for 10 min; less than $\pm 0.25\%$ of setting for 6 hours
Output waveforms	Sine wave, square wave, triangle, ramp, and sine pulse
Sine wave, square wave, and triangle output (selectable):	Amplitude: 0 to 32.5 V p-p open circuit, 90 to 10 V p-p into 50 Ω ; 0 to 30 V p-p into 600 Ω Accuracy: $\pm 1\%$ at 32.5 V p-p

See footnote at end of table.